

	<div data-bbox="1235 218 1474 457"> <p><b>District Impacted</b></p> <p><input type="checkbox"/> 1 = Wolosin</p> <p><input type="checkbox"/> 2 = Woolard</p> <p><input type="checkbox"/> 3 = Scott</p> <p><input type="checkbox"/> 4 = Fowler</p> <p><input type="checkbox"/> 5 = Macaluso</p> <p><input checked="" type="checkbox"/> All</p> </div> <p style="text-align: center;"><b>AGENDA ITEM SUMMARY</b></p>
<b>AGENDA DATE</b>	October 22, 2019
<b>DESCRIPTION</b>	CONSIDER RESOLUTION NO. 2019-R151; A RESOLUTION AUTHORIZING THE CITY MANAGER TO ENTER INTO AND MANAGE AN ANNUAL SERVICE PLAN AGREEMENT BETWEEN THE CITY OF BOERNE AND STRYKER FOR PREVENTATIVE MAINTENANCE AND INSPECTION OF CITY OWNED AUTOMATIC ELECTRONIC DEFIBRILLATORS (AED) AND LIFE PAK UNITS.
<b>STAFF'S RECOMMENDED ACTION (be specific)</b>	Approve Resolution No. 2019-R151; Entering into an annual service plan agreement with Stryker for maintenance of AED's and Life Pak units for a total not to exceed \$16,944.00.
<b>CONTACT PERSON</b>	Doug Meckel, Fire Chief
<b>SUMMARY</b>	<p>The City currently owns forty (40) AED's and one (1) Life Pak 15 monitor / defibrillator. AED's are placed in City facilities and vehicles such as City Hall, library, tennis courts, and baseball fields. The Fire Department provides CPR / AED training which is mandatory for City employees per the City Safety Policy, providing many capable users throughout the City.</p> <p>AED's are expensive to maintain, however, over the years there have been several lives saved in the City by personnel having quick access to such equipment. These units require annual service to be sure they work properly when needed. The service includes checking/replacing the batteries, replacing expired pads, making sure unit will shock properly and is in overall proper working condition. Stryker, the company utilized to provide this service, is proprietary. Stryker has submitted a quote for \$16,944 for a one-year service plan for all forty-one units.</p>
<b>COST</b>	\$16,944
<b>SOURCE OF FUNDS</b>	General Fund
<b>ADDITIONAL INFORMATION</b>	

This summary is not meant to be all inclusive. Supporting documentation is attached.